



Sea-Bird Electronics, Inc.

13431 NE 20th St. Bellevue, Washington 98005 USA
www.seabird.com

Phone: (425) 643-9866

Fax: (425) 643-9954

Email: seabird@seabird.com

Pressure Test Certificate

Test Date: **2017-05-08**

Description: **SBE-43 DO Sensor**

Sensor Information:

Model Number: **SBE-43**

Serial Number: **3582**

Pressure Test Protocol:

Low Pressure Test: **40** PSI Held For: **15** Minutes

High Pressure Test: **870** PSI Held For: **15** Minutes

Passed Test: **True**

Tested By: **RH**

High pressure is generally equal to the maximum depth rating of the instrument

