



Reliable Instruments for an Unreliable World

845 W. Maude Ave. Telephone 408-749-0994
Sunnyvale, CA 94085 Toll Free 877-316-8049
www.turnerdesigns.com Fax 408-749-0998

Specification Certificate

Product: CYCLOPS-7 Submersible Fluorometer, PN 2100-000/2108-000

Application: Ultraviolet: Oil

Note: Oil versions are stamped with the Letter "O" on the connector.

This certificate confirms that the CYCLOPS-7 with **Serial Number:** 2102980 has passed all manufacturing performance testing and meets the specifications detailed below.

For operating instructions, please refer to the CYCLOPS-7 User's Manual contained on the CD disk that is included with the unit or can be viewed at the following web link:
<http://www.turnerdesigns.com/t2/doc/manuals/main.html>

Measurement Parameter	X 1 Gain	X 10 Gain	X 100 Gain
Blank Solution -0.05 Molar H_2SO_4 , millivolt output	0 to 40 mV	0 to 80 mV	0 to 450 mV
Standard deviation in 0.05 Molar H_2SO_4	0.2 mV	2 mV	10 mV
Maximum Concentration value for Quinine Sulfate in 0.05 Molar H_2SO_4	1000 ppb	100 ppb	10 ppb
Gain Factor Spec.	X 1	X 10, +/- 5%	X100, +/- 5%

Note: These specifications are determined from tests performed with Quinine Sulfate in 0.05 Molar H_2SO_4 . There will be some variation in the fluorescence to concentration relationship depending on the preparation of standard solutions.